


Certificate of Quality Control

Manufacturer	SIFIN Institut für Immunpräparate und Nährmedien GmbH Berlin, Berliner Allee 317-321, 13088 Berlin, Germany
Product	Test Serum Anti-Coli O 145:K-
Catalogue No. (Package size)	TS 2221 (1 ml)
Lot No.	86 07 11
Expiry date	2016-07
Intended use	Test Serum Anti-Coli O 145:K- (undiluted) records all E. coli strains that contain the antigen O 145 in slide agglutination. In the Widal reaction the serum reaches a titre of 1:1600 (tube) or 1:3200 (MTP) with the homologous E. coli strains; deviations of ± 1 titre step are acceptable.
Specificity	Since unspecific co-agglutinations with other E. coli strains often occur in the slide agglutination test, a positive result in the slide test must always be confirmed by examination of the boiled E. coli isolate in the Widal test.
Preservatives	Sodium azide 0.9 mg/ ml
Storage	2...8 °C

Berlin, 21.12.11


Head of Quality Control