 SGR11.1 DOC 121 Rev 00 09/04/2014	CERTIFICATE OF STERILITY AND QUALITY ASSURANCE	N° 17-CRT06432
--	---	--------------------------


COPAN CODE	CEPHEID CODE	LOT N°	EXP. DATE
8U003N.PH	URINE/A-50	172484600	31/12/2019

DEVICE DESCRIPTION:	Xpert Urine Specimen Collection Kit A
----------------------------	---------------------------------------

1. Manufacture date: 21/12/2017
2. pH of the medium, acceptability limits pH 7,95 - 8,35 at 20,0-26,0 °C: passed
3. Pipettes have been irradiated by gamma radiation.
The applied dose range is: 25-35 KGY.

All reported results were obtained at the time of lot release.
 Goods have been manufactured according to the ISO 13485 system.
 By signing below, Copan Italia SpA certifies that the above quality requirements have been met and all information stated above is correct.

Made by COPAN ITALIA SpA ,via F. Perotti 10, 25125 Brescia, Italy for CEPHEID



Quality Assurance

Certificate Issue Date
24/01/2018