



## • Certificate of Analysis

Product Name : IVD Bacterial Test Standard

Part No. 8290190

LOT No.: 6030322004

Use by<sup>1)</sup>: 2023-01-31

Date of Production: 2022-02-07

QC release date: 2022-02-17

Storage Conditions : Store at -18 °C or below immediately on arrival.

<u>Test</u>	<u>Specification</u>	<u>LOT Results</u>
Verification of the calibration points by MALDI-TOF MS	Within a mass range of $\pm 900$ ppm	Passed
Relative intensity of reference proteins No. 3 by MALDI-TOF MS	$0.20 \pm 0.05$	Passed
Relative intensity of reference proteins No. 4 by MALDI-TOF MS	$0.22 \pm 0.05$	Passed
MALDI Biotyper identification	Classification Result Report lists "Escherichia coli" as best match.	Passed
	log(score) Value $\geq 2.2$ of the best match	Passed

### Remarks:

- 1) IVD BTS is shipped at ambient temperature but must be stored at  $-18^{\circ}\text{C}$  or below immediately on arrival. The expiry date on the package is valid for the enclosed IVD BTS when stored at  $-18^{\circ}\text{C}$  or below on arrival. Frozen, dissolved Bruker IVD BTS can be stored for up to 5 months at  $-18^{\circ}\text{C}$  or below. See instruction for use at [www.bruker.com/care](http://www.bruker.com/care).

Bruker Daltonics GmbH & Co. KG,  
Germany Fahrenheitstr. 4 – 28359  
Bremen phone +49-421-2205-0

  
Bruker Daltonics GmbH & Co. KG  
Fahrenheitstr. 4  
D-28359 Bremen  
Phone +49 (0)421-2205-0  
Fax +49 (0)421-2205-376  
Dr. Markus Kostrzewa, SAFETY OFFICER