



CERTIFICATE OF ANALYSIS

POSITIVE CONTROL KIT

| | |
|--------------------------------|----------------------------|
| PRODUCT CODE: SI 190.911 | NAME: POSITIVE CONTROL KIT |
| BATCH N.: PCK2002 | EXPIRY DATE: 17/05/2023 |
| MANUFACTURING DATE: 23/03/2022 | |

◆ RECONSTITUTING AND DILUTING SOLUTIONS

| | Physical parameters * | Decontamination at 37 °C after 72 h | Growth inter batch | Growth intra batch |
|-------------------------|-----------------------|-------------------------------------|--------------------|--------------------|
| Eugonic broth | | | | |
| Batch N.: FLA21005 | Expected | Compliant | No growth | CV%≤35% |
| Expire date: 26/11/2023 | Found | OK | OK | 11,7% |

* Volume, transparency, colour

◆ LYOPHILIZED MICROORGANISM

| | Growth inter batch (CV% ≤35%) | Physical check** |
|-------------------------|-------------------------------|------------------|
| Batch N.: 21002 | Expected | Compliant |
| Expire date: 19/10/2023 | Found | OK |

** Aspect

STORAGE CONDITIONS: 2 ÷ 8 °C

Date of Approval:
23/03/2022

Pividori Denis:
Manufacturing Manager

Date of Release:
23/03/2022

We certify that the lot in object has been manufactured and tested according with ALIFAX manufacturing and quality control procedures and specifications. These requirements have been established according to the European Directive 98/79/EC. We confirm that the above referenced product conformed to the specifications and then it is released

Release authorized by Quality Assurance

ALIFAX S.r.l.

Registered and head office
via Petrarca, 2/1 - 35020 Polverara PD Italy
Tel. (+39) 049 0992000 Fax (+39) 049 5855434
info@alifax.com - info.alifax@pec.alifax.com

R&D, Design and Manufacturing Dept
via Merano, 30,33045 Nimis UD Italy
Tel. (+39) 0432 197900 Fax (+39) 0432 547378
info.udine@alifax.com - info.udine@pec.alifax.com

©Copyright 2020. This document is exclusive property of ALIFAX S.r.l, confidential and not freely disclosable. All rights reserved.