

• Certificate of Analysis

Product Name : **IVD Matrix HCCA-portioned**

Part No.: **8290200**

LOT No.: **6020222086**

Use by¹⁾: **2024-09-30**

Date of Production: **2022-10-17**

Storage Conditions : **Store in a refrigerator at 2°C to 8°C on arrival.**

<u>Test</u>	<u>Specification</u>	<u>LOT Results</u>
1. Amount of HCCA content	AQL-value=1.0 Inspection level II U=2.8 mg // L=2.2 mg Sampling in compliance of DIN ISO 3951-1	Passed
2. MALDI Biotyper identification of a 1:20 diluted IVD BTS	Classification Result Report lists " <i>Escherichia coli</i> " as best match with a log score value ≥ 2.000 .	Passed
3. Blank value analysis of IVD Matrix HCCA-portioned by MALDI Biotyper	Classification Result Report lists "No Organism Identification Possible." with a log score value ≤ 1.699 .	Passed

Remarks:

1) IVD Matrix HCCA-portioned is shipped at ambient temperature. The expiry date on the package is valid for the product when stored in a refrigerator at 2°C to 8°C on arrival.

Bruker Daltonics GmbH & CO. KG, Germany
Fahrenheitstr. 4 – 28359 Bremen
phone +49-421-2205-0
QC release date: 2022-10-19



BRUKER
Bruker Daltonics GmbH & Co. KG
Fahrenheitstr. 4
D-28359 Bremen
Phone + 49 (0) 421-2205-0
Fax + 49 (0) 421-2205-370
Dr. Thomas Maier SAFETY OFFICER