

MANUFACTURER:

DIESSSE DIAGNOSTICA SENESE SPA
STRADA DEI LAGHI 39
53035 MONTERIGGIONI (SI),
ITALY

SINGLE REGISTRATION NUMBER

IT-MF-000013311

AUTHORIZED REPRESENTATIVE:

//

PRODUCT:
CODE:**ENZY-WELL/CHORUS WASHING BUFFER**
83606

INTENDED PURPOSE:

To be used in the Enzy-Well immunoenzymatic test kits and
CHORUS instruments as wash solution for the reaction wells.
It must only be used by professional laboratory personnel.

BASIC UDI-DI

803389132CHORUS0028

UDI-DI

08033891322588

RISK CLASS:

CLASS A

CLASSIFICATION RULE:

RULE 5A

CONFORMITY ASSESSMENT ROUTE:

ARTICLE 17, ANNEX II and ANNEX III

WE HEREWITH DECLARE THAT THE ABOVE MENTIONED PRODUCT IS IN CONFORMITY WITH **REGULATION (EU) 2017/746** OF THE EUROPEAN PARLIAMENT AND OF THE COUNCIL ON IN VITRO DIAGNOSTIC MEDICAL DEVICE. ALL SUPPORTING DOCUMENTATION IS RETAINED UNDER THE PREMISES OF THE MANUFACTURER.

REFERENCE TO ANY CS APPLIED:

NOT APPLICABLE

NOTIFIED BODY:

NOT NECESSARY

(EU) CERTIFICATE:

//

REVISION:

0

PLACE, DATE OF ISSUE:

MONTERIGGIONI, 23 MAY 2022

EXPIRY DATE:

//

THE PRESENT EU DECLARATION OF CONFORMITY IS ISSUED UNDER THE SOLE AND EXCLUSIVE RESPONSIBILITY OF THE MANUFACTURER.

SIGNATURE:



MARIA CLAUDIA ALCARO
PERSON RESPONSIBLE FOR THE REGULATORY
COMPLIANCE

THIS IS THE CERTIFIED COPY OF THE ORIGINAL DOCUMENT STORED IN ARCHIVE OF DIESSE DIAGNOSTICA SENESE SPA

ISSUED: MONTERIGGIONI, 2022-06-15

M. Stoczko

MAGDALENA STOCZKO
REGULATORY SUPERVISOR