

• Certificate of Analysis

Product Name : **Bruker Matrix HCCA, portioned**

Part No. 8255344
 LOT No.: 0000241630
 Use by¹⁾: 2018-01
 Date of Production: 2016-02-23

Storage Conditions : Store at 2°C to 8°C on arrival.

<u>Test</u>	<u>Specification</u>	<u>LOT Results</u>
Differential weighing	Tolerance: ± 0.3 mg	Passed
Functional Check against Protein Calibration Standard I (Bruker)	Signal/Noise: ≥ 100	Passed
	Resolution: ≥ 250	Passed

Remarks:

¹⁾ Bruker HCCA is shipped at ambient temperature. The expiry date on the package is valid for the product when stored in a refrigerator at 2°C to 8°C on arrival. Dissolved Bruker HCCA is stable at controlled room temperature (20–25°C) for up to one week.

Bruker Daltonik GmbH, Germany
 Fahrenheitstr. 4 – 28359 Bremen
 phone +49-421-2205-0



B. Schubert
 A. Björn Schubert
 Quality Assurance

QC release date: 24/02/2016

Not for use in diagnostic procedures. For research use only.