

Certificate of Analysis

Product Name: **Bruker Matrix HCCA, portioned**

Part No.: 8255344

LOT No.: 7020121025

Use by¹⁾: 2023-02

Date of Production: 2021-03-10

Storage Conditions : **Store at 2°C to 8°C on arrival.**

<u>Test:</u>	<u>Specification</u>	<u>LOT Results</u>
1. Amount of HCCA content. Sampling in accordance to DIN ISO 3951-1.	AQL-value=1.0 Inspection level II U=2.8 mg // L=2.2 mg	Passed
2. MALDI Biotyper identification of a 1:20 diluted BTS. Sampling in accordance to DIN ISO 2859-1.	Classification Result Report lists "Escherichia coli" as best match. AQL-value=1.0 Inspection level S4 L= 2.000 log(score) of the best match	Passed
3. Blank value analysis of HCCA-portioned by MALDI Biotyper. Sampling in accordance to DIN ISO 2859-1.	Classification Result Report lists "no reliable identification". AQL-value=1.0 Inspection level S4 U= 1.699 log(score)	Passed

Remarks:

- 1) Bruker HCCA is shipped at ambient temperature. The expiry date on the package is valid for the product when stored in a refrigerator at 2°C to 8°C on arrival.

Bruker Daltonik GmbH, Germany
 Fahrenheitstr. 4 – 28359 Bremen
 phone +49-421-2205-0


 Bruker Daltonik GmbH
 Fahrenheitstr. 4
 D-28359 Bremen
 Tel: +49 (0) 421 - 22 05 - 0
 Dr.M. Kostrzewa, SAFETY OFFICER

QC release date: 19-03-2021

Not for use in diagnostic procedures. For research use only.