

• Certificate of Analysis


Product Name: **Bruker Matrix HCCA, portioned**
 Part No.: 8255344
 LOT No.: 6020121023
 Use by¹⁾: 2023-08
 Date of Production: 2021-09-09
 Storage Conditions : **Store at 2°C to 8°C on arrival.**

<u>Test:</u>	<u>Specification</u>	<u>LOT Results</u>
1. Amount of HCCA content. Sampling in accordance to DIN ISO 3951-1.	AQL-value=1.0 Inspection level II U=2.8 mg // L=2.2 mg	Passed
2. MALDI Biotyper identification of a 1:20 diluted BTS. Sampling in accordance to DIN ISO 2859-1.	Classification Result Report lists " <i>Escherichia coli</i> " as best match. AQL-value=1.0 Inspection level S4 L= 2.000 log(score) of the best match	Passed Passed
3. Blank value analysis of HCCA-portioned by MALDI Biotyper. Sampling in accordance to DIN ISO 2859-1.	Classification Result Report lists "no reliable identification". AQL-value=1.0 Inspection level S4 U= 1.699 log(score)	Passed Passed

Remarks:

- 1) Bruker HCCA is shipped at ambient temperature. The expiry date on the package is valid for the product when stored in a refrigerator at 2°C to 8°C on arrival.

Bruker Daltonics GmbH & Co. KG, Germany
 Fahrenheitstr. 4 – 28359 Bremen phone
 +49-421-2205-0



BRUKER
 Bruker Daltonics GmbH & Co. KG
 Fahrenheitstr. 4
 D-28359 Bremen
 Phone + 49 (0)421- 2205-0
 Fax +49 (0)421- 2205-370
 Dr.M. Kostrzewa, SAFETY OFFICER

QC release date: 15-09-2021

Not for use in diagnostic procedures. For research use only.