

• Certificate of Analysis

Product Name : IVD Matrix HCCA-portioned

Part No. 8290200
 LOT No.: 7020220026
 Use by¹⁾: 2022-03
 Date of Production: 2020-04-06
 Storage Conditions : Store in a refrigerator at 2°C to 8°C on arrival.

<u>Test:</u>	<u>Specification</u>	<u>LOT Results</u>
1. Amount of HCCA content	AQL-value=1.0 Inspection level II U=2.8 mg // L=2.2 mg Sampling in compliance of DIN ISO 3951-1	Passed
2. MALDI Biotyper identification of a 1:20 diluted IVD BTS	Classification Result Report lists " <i>Escherichia coli</i> " as best match with a log score value ≥ 2.000 .	Passed
3. Blank value analysis of IVD matrix HCCA-portioned by MALDI Biotyper	Classification Result Report lists "not reliable identification." with a log score value ≤ 1.699 .	Passed

Remarks: ¹⁾ IVD Matrix HCCA-portioned is shipped at ambient temperature. The expiry date on the package is valid for the product when stored in a refrigerator at 2°C to 8°C on arrival.

Bruker Daltonik GmbH, Germany
 Fahrenheitstr. 4 – 28359 Bremen
 phone +49-421-2205-0


 Bruker Daltonik GmbH
 Fahrenheitstraße 4
 D-28359 Bremen
 Dr. M. Kostrzewa, SAFETY OFFICER

QC release date: 2020-04-22