

CERTIFICATE OF ANALYSIS

POSITIVE CONTROL KIT

Rev. 1.1 - 2016.06.22

PRODUCT CODE: SI 190.911	NAME: POSITIVE CONTROL KIT
BATCH N.: PCK17003	
MANUFACTURING DATE: 30/03/2017	EXPIRY DATE: 30/05/2018

◆ *Reconstituting and Diluting Solutions*

Eugonic broth

Batch N.: FLA17002

Expire date:
17/02/2019

	Physical parameters *	Decontamination at 37 °C after 72 h	Growth inter batch	Growth intra batch
Expected	Compliant	No growth	CV%≤35%	CV%≤35%
Found	OK	OK	12.3%	<5%

* Volume, transparency, colour

◆ *Lyophilized Microorganism*

Batch N: 16002

Expire date: 24/11/2019

	Growth inter batch (CV%≤35%)	Physical check**
Expected	Compliant	Compliant
Found	OK	OK

** Aspect

STORAGE CONDITIONS: 2°C – 8 °C

Date of Approval: 03/04/2017

Pividori Denis:
Manufacturing Manager


