

CERTIFICATE OF ANALYSIS

POSITIVE CONTROL KIT

Rev. 1.1 - 2016.06.22

PRODUCT CODE: SI 190.911	NAME: POSITIVE CONTROL KIT
BATCH N.: PCK18005	
MANUFACTURING DATE: 27/08/2018	EXPIRY DATE: 22/12/2019

◆ *Reconstituting and Diluting Solutions*

Eugonic broth

Batch N.: FLA18006

Expire date:
28/09/2020

	Physical parameters *	Decontamination at 37 °C after 72 h	Growth inter batch	Growth intra batch
Expected	Compliant	No growth	CV%≤35%	CV%≤35%
Found	OK	OK	11,1%	<5%

* Volume, transparency, colour

◆ *Lyophilized Microorganism*

Batch N: 18001

Expire date: 15/04/2020

	Growth inter batch (CV%≤35%)	Physical check**
Expected	Compliant	Compliant
Found	OK	OK

** Aspect

STORAGE CONDITIONS: 2°C – 8 °C

Date of Approval: 08/11/2018

Pividori Denis:
Manufacturing Manager


