

CERTIFICATE OF ANALYSIS

POSITIVE CONTROL KIT

Rev. 1.1 - 2016.06.22

PRODUCT CODE: SI 190.911	NAME: POSITIVE CONTROL KIT
BATCH N.: PCK19007	
MANUFACTURING DATE: 02/10/2019	EXPIRY DATE: 25/11/2020

◆ *Reconstituting and Diluting Solutions*

Eugonic broth

Batch N.: FLA19006

Expire date:
26/08/2021

	Physical parameters *	Decontamination at 37 °C after 72 h	Growth inter batch	Growth intra batch
Expected	Compliant	No growth	CV%≤35%	CV%≤35%
Found	OK	OK	15.3%	<5%

* Volume, transparency, colour

◆ *Lyophilized Microorganism*

Batch N: 19001

Expire date:

	Growth inter batch (CV%≤35%)	Physical check**
Expected	Compliant	Compliant
Found	OK	OK

** Aspect

STORAGE CONDITIONS: 2°C – 8 °C

Date of Approval: 08/10/2019

Pividori Denis:
Manufacturing Manager


