



CERTIFICATE OF ANALYSIS

POSITIVE CONTROL KIT

PRODUCT CODE: SI 190.911	NAME: POSITIVE CONTROL KIT
BATCH N.: PCK21004	EXPIRY DATE: 17/06/2022
MANUFACTURING DATE: 26/04/2021	

◆ RECONSTITUTING AND DILUTING SOLUTIONS

Eugonic broth		Physical parameters *	Decontamination at 37 °C after 72 h	Growth inter batch	Growth intra batch
		Expected	Compliant	No growth	CV%≤35%
Batch N.: FLA20005					
Expire date: 16/11/2022		Found	OK	21,8%	<5%

* Volume, transparency, colour

◆ LYOPHILIZED MICROORGANISM

Batch N.: 20002		Growth inter batch (CV% ≤35%)	Physical check**	
		Expected	Compliant	Compliant
Expire date: 17/06/2022		Found	OK	OK

** Aspect

STORAGE CONDITIONS: 2 ÷ 8 °C

Date of Approval:
29/04/2021

Pividori Denis:
Manufacturing Manager

Date of Release:
29/04/2021

We certify that the lot in object has been manufactured and tested according with ALIFAX manufacturing and quality control procedures and specifications. These requirements have been established according to the European Directive 98/79/EC. We confirm that the above referenced product conformed to the specifications and then it is released

Release authorized by Quality Assurance

ALIFAX S.r.l.

Registered and head office
via Petrarca, 2/1 - 35020 Polverara PD Italy
Tel. (+39) 049 0992000 Fax (+39) 049 5855434
info@alifax.com - info.alifax@pec.alifax.com

R&D, Design and Manufacturing Dept
via Merano, 30,33045 Nimis UD Italy
Tel. (+39) 0432 197900 Fax (+39) 0432 547378
info.udine@alifax.com - info.udine@pec.alifax.com

©Copyright 2020. This document is exclusive property of ALIFAX S.r.l, confidential and not freely disclosable. All rights reserved.