

## Certificate of Quality Control

CERTIFIED QUALITY MANAGEMENT SYSTEM IN  
ACCORDANCE WITH DIN EN ISO 9001 AND DIN EN ISO 13485

sifin diagnostics gmbh

Berliner Allee 317-321  
13088 Berlin, Germany

T: +49 30 700 144-0  
F: +49 30 700 144-30  
info@sifin.de  
www.sifin.de

<b>Product</b>	<b>Anti-Coli O 157:K-</b>
Catalogue no.	TR2218
Package size	1 ml
<b>Lot no.</b>	<b>0101119</b>
<b>Expiry date</b>	<b>2024-11</b>
Intended use	Anti-Coli O 157:K- records all <i>E. coli</i> strains that contain the antigen O 157:K- in slide agglutination. In the Widal reaction the serum reaches a titre of 1:1600 (tube, MTP) with the homologous <i>E. coli</i> strains; deviations of $\pm 1$ titre step are acceptable.
Specificity	Since unspecific co-agglutinations with other <i>E. coli</i> strains often occur in the slide agglutination test, a positive result in the slide test must always be confirmed by examination of the boiled <i>E. coli</i> isolate in the Widal test.
Preservatives	Sodium azide 0.9 mg/ ml
Storage	2 °C...8 °C

Berlin, 27.07.2020  
place, date

J. Leuders  
signature Head of Quality Control