

Certificate of Quality Control

CERTIFIED QUALITY MANAGEMENT SYSTEM IN
ACCORDANCE WITH DIN EN ISO 9001 AND DIN EN ISO 13485

sifin diagnostics gmbh

Berliner Allee 317-321
13088 Berlin, Germany

T: +49 30 927 030-0
F: +49 30 927 030-30
info@sifin.de
www.sifin.de

| | |
|--------------------|--|
| Product | Anti-Salmonella H:m |
| Catalogue no. | TS 1413 |
| Package size | 1 ml |
| Lot no. | 1920219 |
| Expiry date | 2024-02 |
| Intended use | Anti-Salmonella H:m records all <i>Salmonella</i> strains, as long as they contain antigen H:m (e. g. H:g,m; H:g,m,s,t; H:m,t) in slide agglutination. |
| Specificity | No agglutination with heterologous <i>Salmonella</i> H antigens and <i>Salmonella</i> O antigens. |
| Preservatives | Sodium azide 0.9 mg/ml |
| Storage | 2 °C...8 °C |

Berlin,
place, date

21.03.2019


signature Head of Quality Control