



## Certificate of Quality Control

CERTIFIED QUALITY MANAGEMENT SYSTEM IN  
ACCORDANCE WITH DIN EN ISO 9001 AND DIN EN ISO 13485


sifin diagnostics gmbh

Berliner Allee 317-321  
13088 Berlin, Germany

T: +49 30 927 030-0  
F: +49 30 927 030-30  
info@sifin.de  
www.sifin.de

<b>Product</b>	<b>Control Serum Anti-Salmonella Typhimurium for the Widal Reaction</b>
Catalogue no.	TS 1624
Package size	1 ml
<b>Lot no.</b>	<b>219 11 14</b>
<b>Expiry date</b>	<b>2019-11</b>
Intended use	The control serum agglutinates in the Widal test with the homologous Salmonella Typhimurium Test Antigen (sifin) with a titre of 1:1600 (tube) and 1:3200 (MTP). Deviations of $\pm$ one titre step are permitted.
Specificity	The serum control (antigen addition is substituted by phosphate buffered saline solution) must not show any agglutination.
Preservatives	Sodium azide 0.9 mg/ml
Storage	2 °C...8 °C

Berlin, 18.11.14  
place, date

  
signature Head of Quality Control