

## Certificate of Quality Control

CERTIFIED QUALITY MANAGEMENT SYSTEM IN  
ACCORDANCE WITH DIN EN ISO 9001 AND DIN EN ISO 13485

<b>Product</b>	<b>Anti-Shigella flexneri</b> type 1-6, Gr.3,4 (y), 6, 7,8 (x)
Catalogue no.	TS1901
Package size	1 ml
<b>Lot no.</b>	<b>0210121</b>
<b>Expiry date</b>	<b>2025-12</b>
Intended use	Anti-Shigella flexneri records all type- and group-specific antigens of <i>Shigella flexneri</i> in slide agglutination.
Specificity	No agglutination with the serotypes of <i>Shigella dysenteriae</i> (type 1-10), <i>Shigella boydii</i> (type 1 to 15) and <i>Shigella sonnei</i> (Phase I and Phase II).
Preservatives	Sodium azide 0.9 mg/ml
Storage	2 °C...8 °C

Berlin, 07.06.27  
\_\_\_\_\_  
place, date

7. Leuders  
\_\_\_\_\_  
signature Head of Quality Control