

Certificate of Quality Control

CERTIFIED QUALITY MANAGEMENT SYSTEM IN ACCORDANCE WITH DIN EN ISO 9001 AND DIN EN ISO 13485

sifin diagnostics gmbh

Berliner Allee 317-321 13088 Berlin, Germany

T: +49 30 700 144-0 F: +49 30 700 144-30 info@sifin.de www.sifin.de

Product

Anti-Coli O 103:K-

Catalogue no.

TS2216

Package size

1 ml

Lot no.

7651021

Expiry date

2026-09

Intended use

Anti-Coli O 103:K- records all E. coli strains that contain the antigen

O 103:K- in slide agglutination. In the Widal reaction the serum reaches a titre

of 1:3200 (tube) and 1:6400 (MTP) with the homologous E. coli strains;

deviations of ±1 titre step are acceptable.

Specificity

Since unspecific co-agglutinations with other *E. coli* strains often occur in the slide agglutination test, a positive result in the slide test must always be confirmed by examination of the boiled *E. coli* isolate in the Widal test.

Preservatives

Sodium azide 0.9 mg/ ml

Storage

2 °C...8 °C

Berlin, 16.11.27

place, date

signature Head of Quality Control

J. Kirch