

CERTIFICATE OF ANALYSIS

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and testing specifications required by Cepheid's Quality System, in compliance with the US Food and Drug Administration's Quality System Requirements, ISO 13485, European IVD Directive and the Canadian Medical Devices Regulations (CMDR).

		(3.12.11).		
Product Name:	Xpert® Xpress Flu/RSV			
Cepheid Catalogue F	Part No.: XPRSFLU/RSV-CE-10			
Kit Lot No.: 1000	0147067			
Cartridge Lot No.: 10	0708			
Kit Expiration Date:	2020-07-05			
Legal Manufacturer		Manufacturing Facility		
Cepheid 904 Caribbean Drive Sunnyvale, CA 94089 U	JSA	Cepheid AB Röntgenvägen 5 SE-171 54 Solna Sweden	,	
Functional Testing				
Test Description	Acceptance Criteria		Test Result	
Negative	Flu A NEGATIVE; Flu B NEGATIVE; RSV NEGATIVE		Passed	
Low Positive	Flu A POSITIVE; Flu B POSITIVE; RSV POSITIVE		Passed	

Flu A POSITIVE; Flu B POSITIVE; RSV POSITIVE

Signature of Quality Assurance Date

Name: Marie Bah

High Positive

Title: QA Manager Release

Passed