



CERTIFICATE OF ANALYSIS

ESR_COA_SI305#2270B.pdf

Product Code SI 305.300-A	Name LATEX CONTROLS	Quantity manufactured 178	Type of Tubes Greiner
-------------------------------------	-------------------------------	-------------------------------------	---------------------------------

Lot Number 2270B	Manufacturing Date 17/02/2022	Expiry Date 16/08/2022		
----------------------------	---	----------------------------------	--	--

Standard Values			Measured Values				Internal	Volume
	Transmittance	ESR (mm/h) (min-Val-max)	ESR (mm/h) (min-Val-max)	Instrument	Serial Number	OK	Aggregation	checked
Level 2	1534	6 - 11	9 9 9	TEST 1 BCL	2173	<input checked="" type="checkbox"/>	NO	YES
Level 3	980	15 - 22	19 20 20	TEST1 THL	T13290	<input checked="" type="checkbox"/>	NO	YES
Level 4	637	56 - 74	65 67 68	ROLLER 20 PN	133	<input checked="" type="checkbox"/>	NO	YES

Controlled By **Cirio Alberto**
Approved By Manufacturing Manager **Pividori Denis**

Date of Control **18/02/2022**
Date of Approval **18/02/2022**

We certify that the lot in object has been manufactured and tested according with ALIFAX manufacturing and quality control procedures and specifications. These requirements have been established according to the European Directive 98/79/EC. We confirm that the above referenced product conformed to the specifications and then it is released.

Released By Quality **Zorzut Milena**

Date of Release **18/02/2022**

This is a computer-generated document. No signature is required.

Registered Office

Via Petrarca, 2/1 - 35020 Polverara (PD) Italy
Tel (+39) 049 0992000 - Fax (+39) 049 5855434
Mail PEC: info.alifax@pec.alifax.com
info@alifax.com - www.alifax.com



Manufacturing Site

Via Merano, 30 - 33045 Nimis (UD) Italy
Tel (+39) 0432 547454 - Fax (+39) 0432 547378
Mail PEC: info.udine@pec.alifax.com
info.udine@alifax.com