

CERTIFICATE OF ANALYSIS

Product Name: Streptococcal Grouping Reagent Latex Suspension Group B	Lot Number: C20227A	
Product Code: PL.032	Expiration Date: 2023-08-31	


Additional Information

Storage Conditions:	2-8°C
Precautions:	Refer to Instructions for Use

Quality Control Testing

Appearance	Reaction	Result
Homogenous suspension with no agglutination	N/A	Satisfactory
Organisms		
<i>Streptococcus agalactiae</i> ATCC 12386 (Group B)	Positive	PASS
<i>Streptococcus agalactiae</i> ATCC 13813 (Group B)	Positive	PASS
<i>Streptococcus</i> Group B - Clinical Isolates	Positive	PASS
<i>Streptococcus pyogenes</i> ATCC 19615 (Group A)	Negative	PASS
<i>Streptococcus equisimilis</i> ATCC 12388 (Group C)	Negative	PASS
<i>Enterococcus faecalis</i> ATCC 49477 (Group D)	Negative	PASS
<i>Streptococcus</i> sp. ATCC 12392 (Group F, type 2)	Negative	PASS
<i>Streptococcus</i> sp. ATCC 12394 (Group G)	Negative	PASS
<i>Listeria monocytogenes</i> NCTC 11994	Negative	PASS
Control		
Polyvalent Positive Control (PL.040)	Positive	PASS

Release

2022 04 04	
Date	Quality Control

Canada ☎ 800 268 2341 Fax 905 731 0206
20 Mural Street, Unit #4, Richmond Hill, Ontario,
L4B 1K3
support@pro-lab.com

U.S.A. ☎ (512) 832 9145 Fax (512) 832 6424
21 Cypress Blvd., Suite 1155, Round Rock,
Texas 78665-1046
support@pro-lab.us

U.K. ☎ 0151 353 1613 Fax 0151 353 1614
3 Bassendale Road, Bromborough, Wirral,
Merseyside CH62 3QL
uksupport@pro-lab.com