

## QUALITY CONTROL CERTIFICATE

### **NYSTATIN 50 µg Neo-Sensitabs**

REF: 82212

LOT: 220020

Date of analysis: 2022-07

Date of expiry: 2027-07

Appearance: Light yellow, plane tablet, 9 mm in diameter,  
on both sides printed with a black identification  
code.  
Complies.

Thickness: Mean of 10: 1.50-1.65 mm.  
Complies.

Performance:

Test strains:

*Candida albicans* ATCC 64548

*Candida albicans* ATCC 10231

Product complies with specifications.

Uniformity: Inhibition zones of individual tablets  
vary less than  $\pm 1$  mm from mean.  
Complies.

Microbiological quality: Complies.

Rosco Diagnostica A/S



Peter Samuelsen  
Manager Quality Assurance