

## **CERTIFICATE OF ANALYSIS**

ESR\_COA\_SI305#2358B.pdf

Product Code SI 305.300-A	Name  LATEX CONTROLS	Quantity manufactured <b>178</b>	Type of Tubes  Greiner
Lot Number 2358B	Manufacturing Date Expiry Date  15/12/2022 13/06/2023		

Standard Values					ECD (many /h)			Measured				
	Transmittance	ESR (mm/h) (min-Val-max)		ESR (mm/h) (min-Val-max)			Instrument	Serial Number	OK	Aggregation	checked	
Level 2	1534	6	-	11	9	9	10	TEST1 THL	6608	X	NO	YES
Level 3	980	15	-	22	19	20	20	TEST1 THL	T13290	X	NO	YES
Level 4	637	56	-	<i>74</i>	65	67	68	ROLLER 20 PN	133	X	NO	YES

We certify that the lot in object has been manufactured and tested according with ALIFAX manufacturing and quality control procedures and specifications. We confirm that the above referenced product conformed to the specifications and then it is approved.

Approved by Quality Assurance ...... Pontoni Laura Date of Release 16/12/2022

This is a computer-generated document. No signature is required.

**Registered Office** 

Via Petrarca, 2/1 - 35020 Polverara (PD) Italy Tel (+39) 049 0992000 - Fax (+39) 049 5855434 Mail PEC: info.alifax@pec.alifax.com info@alifax.com - www.alifax.com



**Manufacturing Site** 

Via Merano, 30 - 33045 Nimis (UD) Italy Tel (+39) 0432 547454 - Fax (+39) 0432 547378 Mail PEC: info.udine@pec.alifax.com info.udine@alifax.com