

Certificate of Quality Control

CERTIFIED QUALITY MANAGEMENT SYSTEM IN
ACCORDANCE WITH DIN EN ISO 9001 AND DIN EN ISO 13485

Product	Anti-Salmonella O:9
Catalogue no.	TR1307
Package size	1 ml
Lot no.	6260822
Expiry date	2027-07
Intended use	Anti-Salmonella O:9 records all <i>Salmonella</i> strains, as long as they contain antigen O:9 (group D) in slide agglutination.
Specificity	No agglutination with heterologous <i>Salmonella</i> O antigens and <i>Salmonella</i> H antigens.
Preservatives	Sodium azide 0.9 mg/ml
Storage	2 °C...8 °C

Berlin,
place, date

20.09.2022


signature Head of Quality Control