

## **CERTIFICATE OF ANALYSIS**

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and testing specifications required by Cepheid's Quality System, in compliance with the US Food and Drug Administration's Quality System Requirements, ISO 13485, European IVD Directive and the Canadian Devices Regulations.

IVD Directive and the Can	adian Devices Regulations.	European
Product Name: Xpert® Xpress Flu/RSV		
Cepheid Catalogue Part I	No.: XPRSFLU/RSV-CE-10	
Kit Lot No.: 1000566515		
Cartridge Lot No.: 15406		
Kit Expiration Date: 2024-10-27		
Legal Manufacturer	Manufacturing Facility Solna Sunnyvale	
Cepheid 904 Caribbean Drive Sunnyvale, CA 94089 USA	Cepheid AB  Röntgenvägen 5  SE-171 54 Solna  Sweden	di
Functional Testing		
Test Description	Acceptance Criteria	Test Result
Negative	Flu A NEGATIVE;Flu B NEGATIVE;RSV NEGATIVE	Passed
Low Positive	Flu A POSITIVE;Flu B POSITIVE;RSV POSITIVE	Passed
High Positive	Flu A POSITIVE;Flu B POSITIVE;RSV POSITIVE	Passed
Les 202	is produced electronically and valid without a wet signature $9 - 11 - 13$	are.
Signature of Quality Assu  Name: Eleni Vangeli	rance, Date	
SEES AND TO THE RESONANT SEES SEES AND THE S		
Title: QA Analyst		