

CERTIFICATE OF ANALYSIS

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and testing specifications required by Cepheid's Quality System, in compliance with the US Food and Drug Administration's Quality System Requirements, ISO 13485, European IVD Directive and the Canadian Medical Device Regulations (CMDR).

Product Name: Xpert® HBV Viral Load				
Cepheid Catalogue Part No.: GXHBV-VL-CE-10				
Kit Lot No.: 10001	176782			
Cartridge Lot No.: 13202				
Kit Expiration Date: 2	020-10-04			
Legal Manufacturer		Manufacturing Facility	Solna	Sunnyvale
Cepheid AB Röntgenvägen 5 SE-171 54 Solna Sweden		Cepheid AB Röntgenvägen 5 SE-171 54 Solna Sweden		
Functional Testing				
Test Description	Acceptance Criteria			Test Result
LOW Positive	HBV DETECTED: 1.26-1.70 log10 IU/mL			Passed
HIGH Positive	HBV DETECTED: 5.54-5.86 log10 IU/mL			Passed
Negative	HBV NOT DETECTED			Passed
Kambie Cheala 20191125				
Signature of Quality Assurance Date				
Name: Kambiz Ghaderi				
Title: QA Analyst				