


CERTIFICATE OF ANALYSIS

Product Name: Staph Latex Kit		Lot Number: C21364
Product Code: PL.080B		Expiration Date: 2024-08-31
Product Code	Quantity	Components
PL.083B	2 x 2.5 ml	Staph Latex Reagent
PL.085B	1 x 2.5 ml	Negative Control Latex Reagent
	200	Mixing Sticks
PL.092-20	20 cards	Test Cards
Additional Information		
Storage Conditions:		2-8°C
Instructions for Use:		Refer to pro-lab.com/resources for current version
Quality Control Testing		
Criteria		Result
Strong agglutination of Staph Latex Reagent when tested against <i>Staphylococcus aureus</i> .		PASS
No agglutination of Negative Control when tested against test isolates.		PASS
Release		
2023-03-14		
Date	Quality Control	

Canada ☎ 800 268 2341 Fax 905 731 0206
20 Mural Street, Unit #4, Richmond Hill, Ontario,
L4B 1K3
support@pro-lab.com

U.S.A. ☎ (512) 832 9145 Fax (512) 832 6424
21 Cypress Blvd., Suite 1155, Round Rock,
Texas 78665-1046
support@pro-lab.us

U.K. ☎ 0151 353 1613 Fax 0151 353 1614
3 Bassendale Road, Bromborough, Wirral,
Merseyside CH62 3QL
uksupport@pro-lab.co.uk