

CERTIFICATE OF ANALYSIS

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and testing specifications required by Cepheid's Quality System, in compliance with the US Food and Drug Administration's Quality System Requirements, ISO 13485, European IVD Directive and the Canadian Devices Regulations.

Product Name: Xpert® Xpress CoV-2/Flu/RSV plus						
Cepheid Catalogue Pa	art No.: XP3CO	V2/FLU/RSV-10				
Kit Lot No.: 10007111	51					
Cartridge Lot No.: 15	121					
Kit Expiration Date: 2	2024-01-21					
Legal Manufacturer	Manufactur	ring Facility	Solna	Sunnyvale		
Cepheid	Cepheid AB		Newark	OLodi IVD (B2)		
904 Caribbean Drive	Röntgenväge	en 5				
Sunnyvale, CA 94089	SE-171 54 S	olna				
USA	Sweden					

Functional T	Testing
--------------	---------

Test Description	Acceptance Criteria	Test Result	
Negative	SARS-CoV-2 NEGATIVE;Flu A NEGATIVE;Flu B NEGATIVE;RSV NEGATIVE	Passed	
Positive	SARS-CoV-2 POSITIVE;Flu A POSITIVE;Flu B POSITIVE;RSV POSITIVE	Passed	

■If checked this document is produced electronically and valid without a wet signature.

			ity Assu		20	23- Date	-02-	-C
Name:	Anto	n Ande	ersson En	gström	1			
Title:	QA A	nalyst						