


CERTIFICATE OF ANALYSIS

| | | |
|---|--|--|
| Product Name: Prolex™ Staph Xtra Latex Kit | | Lot Number: C21366 |
| Product Code: PL.1080 | | Expiration Date: 2024-09-30 |
| Product Code | Quantity | Components |
| PL.1083 | 2 x 2.5 ml | Staph Xtra Latex Reagent |
| PL.1085 | 1 x 5.0 ml | Staph Xtra Negative Control Latex Reagent |
| | 200 | Mixing Sticks |
| | 20 cards | Test Cards |
| Additional Information | | |
| Storage Conditions: | | 2-8°C |
| Instructions for Use: | | Refer to pro-lab.com/resources for current version |
| Quality Control Testing | | |
| Criteria | | Result |
| Strong agglutination of Staph Xtra Latex Reagent when tested against <i>Staphylococcus aureus</i> and MRSA* including capsular types 5 and 8 within 20 seconds. | | PASS |
| No agglutination of Staph Xtra Negative Control when tested against test isolates at 20 seconds. | | PASS |
| * MRSA - (Methicillin-resistant <i>Staphylococcus aureus</i>) | | |
| Release | | |
| 2023-03-08 |  | |
| Date | Quality Control | |

Canada ☎ 800 268 2341 Fax 905 731 0206
 20 Mural Street, Unit #4, Richmond Hill, Ontario,
 L4B 1K3
 support@pro-lab.com

U.S.A. ☎ (512) 832 9145 Fax (512) 832 6424
 21 Cypress Blvd., Suite 1155, Round Rock,
 Texas 78665-1046
 support@pro-lab.us

U.K. ☎ 0151 353 1613 Fax 0151 353 1614
 3 Bassendale Road, Bromborough, Wirral,
 Merseyside CH62 3QL
 uksupport@pro-lab.co.uk