

CERTIFICATE OF ANALYSIS

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and testing specifications required by Cepheid's Quality System, in compliance with the US Food and Drug Administration's Quality System Requirements, ISO 13485, European IVD Directive and the Canadian Medical Device Regulations (CMDR).

Product Name: Xpert® TV		
Cepheid Catalogue Part No.: GXTV-CE-	10	
Kit Lot No.: 1000515092		
Cartridge Lot No.: 20306		
Kit Expiration Date: 2024-03-10		
Legal Manufacturer	Manufacturing Facility	Sunnyvale
Cepheid	Cepheid	•
904 Caribbean Drive	904 Caribbean Drive	
Sunnyvale, CA 94089 USA	Sunnyvale, CA 94089 USA	
	8 - 4	
Functional Testing		
Test Description	Acceptance Criteria	Test Result
LOW Positive	TV detected	
HIGH Positive	TV detected	

TV not detected

Signature of Quality Assurance Date

Name: Molly Doan

Negative

Title: Quality Systems Specialist

Passed