

CERTIFICATE OF ANALYSIS

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and testing specifications required by Cepheid's Quality System, in compliance with the US Food and Drug Administration's Quality System Requirements, ISO 13485, European IVD Directive and the Canadian Medical Device Regulations (CMDR).

Product Name:	Xpert® MRSA NXG		
Cepheid Catalog	gue Part No.: GXMRS	A-NXG-C	CE-10
Kit Lot No.:	1000163437	e e	
Cartridge Lot N	o.: 05110		
Kit Expiration I	Date: 2021-07-18		

Legal Manufacturer
Cepheid
904 Caribbean Drive
Sunnyvale, CA 94089 USA

Manufacturing Facility Solna Sunnyvale
Cepheid
904 Caribbean Drive
Sunnyvale, CA 94089 USA

Functional Testing

Test Description	Acceptance Criteria	Test Result
LOW Positive	MRSA detected	Passed
HIGH Positive	MRSA detected	Passed
Negative	MRSA not detected	Passed

2	Equants	28 AUG 1	19
Signati	ure of Quality Assurance	Date	k,
Name	Emaela S. Quan-Tong		
Title:	Quality Systems Specialist	- ·	