


DECLARATION OF CONFORMITY



Product name / Trade name		REF IS-4500NG	BASIC UDI-DI: 5060169696457YN
IDS ACTH II		GMDN: 54055	EMDN: W0102060401
Intended Purpose	For In Vitro Diagnostic Use For Laboratory Professional Use IDS ACTH II assay is an automated <i>in vitro</i> diagnostic device intended for the quantitative, determination of ACTH in human EDTA plasma on the IDS system. Results are to be used in conjunction with other clinical and laboratory data as an aid in the assessment of pituitary and adrenal gland function and the differential diagnosis of hyper- and hypo-cortisolism.		
		EC	REP
Immunodiagnostic Systems Limited, 10 Didcot Way, Boldon Business Park, Boldon, Tyne and Wear, NE35 9PD, UK		Immunodiagnostic Systems SA 101 Rue Ernest Solvay B-4000 Liege Belgium	
SINGLE REGISTRATION NUMBER: GB-MF-000015851		SINGLE REGISTRATION NUMBER: BE-AR-000015342	

RISK CLASS:☐ A ☒ B ☐ C ☐ D**CLASSIFICATION RULE (ANNEX VIII) :** Rule 6**CONFORMITY ROUTE:**☒ ANNEX IX Full Quality System
(Class B, C & D)☐ ANNEX I & II+III
(non-sterile Class A)**Name of the Notified Body / Identification:**TÜV Rheinland LGA Products GmbH.
0197**CE Marking Date:**30th October 2023**EU CERTIFICATE No.:**

HX 2333083-1

COMMON SPECIFICATIONS:

Not Applicable

We, as the manufacturer of the device(s) take sole responsibility for and hereby declare that the above-mentioned product(s) meet(s) the provisions of the following Regulation(s)/Directives

- Regulation EU 2017/746 on *In vitro* Diagnostic Medical Devices

Date: 15th April 2024

Signed on behalf of Immunodiagnostic Systems Limited

Place: UK


M Henderson
RA Manager & PRRCDate: 15th April 2024

Signed on behalf of Immunodiagnostic Systems Limited

Place: UK


D. Mullington
Group QA/RA Director & PRRC