

CERTIFICATE OF ANALYSIS

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and testing specifications required by Cepheid's Quality System, in compliance with the US Food and Drug Administration's Quality System Requirements, ISO 13485, European IVD Directive and the Canadian Devices Regulations.

Product Name:	Xpert® Xpr	ess CoV-2/Flu	/RSV plus		
Cepheid Catalogu	e Part No.:	XP3COV2/I	FLU/RSV-10		
Kit Lot No.: 10015	06041				
Cartridge Lot No.:	33201				
Kit Expiration Dat	e: 2026-10-18				
Legal Manufacture	<u>er</u> <u>N</u>	Ianufacturing	<u>Facility</u>	Solna	Sunnyvale
Cepheid 904 Caribbean Drive Sunnyvale, CA 9408	F	Cepheid AB Röntgenvägen 5 SE-171 54 Soln		Newark	Lodi IVD (B2)

Functional Testing

USA

Test Description	Acceptance Criteria	Test Result
Negative	SARS-CoV-2 NEGATIVE;Flu A NEGATIVE;Flu B NEGATIVE;RSV NEGATIVE	Passed
Positive	SARS-CoV-2 POSITIVE;Flu A POSITIVE;Flu B POSITIVE;RSV POSITIVE	Passed

☑ If checked this document is produced electronically and valid without a wet signature. *Lava Maroof*

Sweden

Signati	ure of Quality Assurance,	Date
Name:	Lava Maroof	
Title:	QA Analyst	