



CERTIFICATE OF ANALYSIS

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and testing specifications required by Cepheid's Quality System, in compliance with the US Food and Drug Administration's Quality System Requirements, ISO 13485, European IVD Directive and the Canadian Medical Device Regulations (CMDR).

Product Name: Xpert® HCV Viral Load

Cepheid Catalogue Part No.: GXHCV-VL-CE-10

Kit Lot No.: 1001512140

Cartridge Lot No.: 52504

Kit Expiration Date: 2027-03-21

Legal Manufacturer

Cepheid AB
Röntgenvägen 5
SE-17154 Solna
Sweden

Manufacturing Facility

Cepheid AB
Röntgenvägen 5
SE-171 54 Solna
Sweden

Solna Sunnyvale

Functional Testing according to D51112 Rev: G

| <i>Test Description</i> | <i>Acceptance Criteria</i> | <i>Test Result</i> |
|--------------------------------|---|---------------------------|
| LOW Positive | HCV DETECTED: 1.50-1.90 log ₁₀ IU/mL | Passed |
| HIGH Positive | HCV DETECTED: 4.25-4.55 log ₁₀ IU/mL | Passed |
| Negative | HCV NOT DETECTED | Passed |

If checked this document is produced electronically and valid without a wet signature.

Lava Maroof

Lava Maroof (12/18/2025 09:33:06 GMT+1)

Signature of Quality Assurance,

Date

Name: Lava Maroof

Title: QA Analyst