

CERTIFICATE OF ANALYSIS

Product Name: Streptococcal Grouping - Extraction Reagent 2	Lot Number: C1383F	
Product Code: PL.038	Expiration Date: 2027-04-30	

Additional Information

Storage Conditions:	2-8°C
Instructions for Use:	Refer to pro-lab.com/resources for current version

Quality Control Testing

Organism	Reaction*	Result
<i>Streptococcus pyogenes</i> ATCC 19615 (Group A)	Positive	PASS
<i>Streptococcus agalactiae</i> ATCC 12386 (Group B)	Positive	PASS
<i>Streptococcus equisimilis</i> ATCC 12388 (Group C)	Positive	PASS
<i>Enterococcus faecalis</i> ATCC 49477 (Group D)	Positive	PASS
<i>Streptococcus</i> sp. ATCC 12392 (Group F)	Positive	PASS
<i>Streptococcus</i> sp. ATCC 12394 (Group G)	Positive	PASS

*When used in conjunction with Extraction Reagents 1 and 3 to extract group specific carbohydrate antigens and tested against homologous latex reagents.

Release

2025-06-19	
Date	Quality Control

Pro-Lab Diagnostics U.K.
3 Bassendale Road, Bromborough, Wirral,
Merseyside CH62 3QL
United Kingdom
Tel.: 0151 353 1613
www.pro-lab.co.uk
E-mail: uksupport@pro-lab.co.uk

Pro-Lab Diagnostics Canada
20 Mural Street, Unit #4
Richmond Hill, Ontario, L4B 1K3
Canada
Tel.: 905-731-0300
www.pro-lab.com
E-mail: support@pro-lab.com

Pro-Lab Diagnostics U.S.A.
1301 Blue Ridge Drive, Suite #101
Georgetown, Texas 78626-3269
United States
Tel.: 512-832-9145
www.pro-lab-direct.com
E-mail: support@pro-lab.us