

## **CERTIFICATE OF ANALYSIS**

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and testing specifications required by Cepheid's Quality System, in compliance with the US Food and Drug Administration's Quality System Requirements, ISO 13485, European IVD Directive and the Canadian Medical Device Regulations (CMDR).

**Product Name:** Xpert® HCV Viral Load

**Cepheid Catalogue Part No.:** GXHCV-VL-CE-10

**Kit Lot No.:** 1001524062

**Cartridge Lot No.:** 53401

**Kit Expiration Date:** 2027-08-08

**Legal Manufacturer**

Cepheid AB  
Röntgenvägen 5  
SE-17154 Solna  
Sweden

**Manufacturing Facility**

Cepheid AB  
Röntgenvägen 5  
SE-171 54 Solna  
Sweden

☒ Solna ☐ Sunnyvale

***Functional Testing according to D51112 Rev: G***

| <b><i>Test Description</i></b> | <b><i>Acceptance Criteria</i></b>               | <b><i>Test Result</i></b> |
|--------------------------------|-------------------------------------------------|---------------------------|
| LOW Positive                   | HCV DETECTED: 1.50-1.90 log <sub>10</sub> IU/mL | Passed                    |
| HIGH Positive                  | HCV DETECTED: 4.25-4.55 log <sub>10</sub> IU/mL | Passed                    |
| Negative                       | HCV NOT DETECTED                                | Passed                    |

☒ If checked this document is produced electronically and valid without a wet signature.

*Lava Maroof*

Lava Maroof (03/20/2026 12:12:12 GMT+1)

**Signature of Quality Assurance,**

**Date**

**Name:** Lava Maroof

**Title:** QA Analyst