Certificate of Quality Control

CERTIFIED QUALITY MANAGEMENT SYSTEM IN
ACCORDANCE WITH DIN EN ISO 9001 AND DIN EN ISO 13485

sifin diagnostics gmbh

Berliner Allee 317-321
13088 Berlin, Germany

T: +49 30 700 144-0
F: +49 30 700 144-30
info@sifin.de
www.sifin.de

Product	**Anti-Shigella sonnei S- and F-form (phase I and II)**
Catalogue no.	TR2014
Package size	1 ml
Lot no.	**6200925**
Expiry date	**2030-08**
Intended use	Anti-Shigella sonnei S- and F-form (phase I and II) records the strains of *Shigella sonnei* S- (phase I) and F-form (phase II) in slide agglutination.
Specificity	No agglutination with antigens of *Shigella flexneri*, *Shigella dysenteriae* (type 1 to 10) and *Shigella boydii* (type 1 to 15).
Preservatives	Sodium azide 0.9 mg / ml
Storage	2 °C…8 °C

This document has been established electronically and is valid without signature.
The batch release was done in accordance with SOP QMV 7.