

• Certificate of Analysis

Product Name : **MBT STAR-Carba IVD Kit**

Part No. **1848467**
 LOT No.: **0000368383**
 Use by: **2021-12**
 Date of Production: **2021-02-25**

Storage Conditions : The MBT STAR-Carba IVD Kit is shipped at ambient temperature but must be stored at **-18 to -25°C immediately on arrival.**

<u>Test</u>	<u>Specification</u>	<u>LOT Results</u>
Instrument calibration for MBT STAR-BL Assay mass range using MBT Compass IVD and MBT STAR-BL IVD Module.	Successfully automated MBT STAR Calibrator based instrument calibration:	
	Sample 1	Passed
	Sample 2	Passed
	Sample 3	Passed
	Sample 4	Passed
	Sample 5	Passed
MBT STAR-BL Assay Performance Test using MBT STAR-Carba IVD Kit according to the IFU	MBT STAR-BL IVD Module Results:	
	1. negative control (- ctrl) as non-hydrolyzed	Passed
	2. positive control (+ ctrl; KPC) as hydrolyzed	Passed
	3. negative QC Sample (- QC) as non-hydrolyzed	Passed
	4. positive QC Sample (+ QC; NDM-1) as hydrolyzed Passed	Passed

Remarks:

Kit components have to be handled carefully to avoid contaminations.

If stored unopened at -18 to -25°C the product is stable until expiration date.

Bruker Daltonik GmbH, Germany
 Fahrenheitstr. 4 – 28359 Bremen
 phone +49-421-2205-0



Dr. Markus Kostrzewa, SAFETY OFFICER

For In Vitro Diagnostic Use.