

Certificate of Analysis

PRODUCT DESCRIPTION: NADAL® COVID-19 Ag Test (CE0197)

PRODUCT CODE	LOT NUMBER	LOT AMOUNT	EXPIRATION DATE	PHYSICAL APPEARANCE
243117N-05	178034	100000PCS	2023-10	Conforms

PRODUCT TEST RESULTS

TEST	QC CRITERION	RESULTS	PASS (Y/N)
Negative Specimen	Negative	Negative	Yes
COV Low Positive	Positive	Positive	Yes
COV High Positive	Positive	Positive	Yes

PRODUCT MANUFACTURING:

No radioactive material of any kind is utilized in the product, or in the manufacturing of the product. The product specified above does not contain any virus, reagent by-product of the same, or metabolic by-product of Hepatitis A, B, C, or D.

If this product is shipped as bulk, nal von minden GmbH assumes no responsibility for the packaging and label claims.

Accepted:

Sandra von Minden

Qualitätsmanagement
nal von minden GmbH



nal von minden GmbH
Carl-Zeiss-Straße 12
47445 Moers
Tel: +49 2841 99820-0
Fax: +49 2841 99820-1
info@nal-vonminden.de

Moers, 2021-11-29

nal von minden GmbH

Carl-Zeiss-Straße 12
47445 Moers
Deutschland

Friedenstraße 32
93053 Regensburg
Deutschland

Tel.: +49 941 29010-0
Fax: +49 941 29010-50

info@nal-vonminden.com
www.nal-vonminden.com

Geschäftsführer:
Sandra von Minden
Roland Meißner
Thomas Zander

Registergericht Kleve:
HRB 5679
Steuer-Nr. 244/133/00130
UST-ID-Nr. DE 189 016 086

Sparkasse Regensburg
Kto: 840 017 073
BLZ: 750 500 00

Deutsche Bank
Kto: 056 998 800
BLZ: 350 700 24

Schnelltests

Labordiagnostika

Laborservice

Beratung & Service