

QUALITY CONTROL CERTIFICATE

AMPHOTERICIN B 10 µg Neo-Sensitabs

REF: 81012

LOT: 1601-1

Date of analysis: 2016-07

Date of expiry: 2019-07

Appearance: Light yellow, plane tablet, 9 mm in diameter,
on both sides printed with a black identification
code.
Complies.

Thickness: Mean of 10: 1.50-1.65 mm.
Complies.

Performance:

Test strains:

Candida albicans ATCC 90028
Candida parapsolis ATCC 22019
Candida krusei ATCC 6258


Product complies with specifications.

Uniformity: Inhibition zones of individual tablets
vary less than ± 1 mm from mean.
Complies.

Microbiological quality: Complies.

Manager Quality Assurance

Rosco Diagnostica A/S



Michela Venturi
Manager Quality Assurance