

## QUALITY CONTROL CERTIFICATE

### **CLOTRIMAZOLE 10 µg Neo-Sensitabs**

REF: 81212N

LOT: 1301-1

Date of analysis: 09-2013

Date of expiry: 09-2018

Appearance: Light yellow, plane tablet, 9 mm in diameter,  
on both sides printed with a black identification  
code.  
Complies.

Thickness: Mean of 10: 1.50-1.65 mm.  
Complies.

Performance:

Test strains:

*Candida albicans* ATCC 10231  
*Candida pseudotropicalis* MIHE 2337

Product complies with specifications.

Uniformity: Inhibition zones of individual tablets  
vary less than  $\pm 1$  mm from mean.  
Complies.

Microbiological quality: Complies.

ROSCO DIAGNOSTICA A/S



José B. Casals  
Quality Assurance Manager