

QUALITY CONTROL CERTIFICATE

CLOTRIMAZOLE 10 µg Neo-Sensitabs

REF: 81212

LOT: 2001-1

Date of analysis: 2020-07

Date of expiry: 2025-06

Appearance: Light yellow, plane tablet, 9 mm in diameter,
on both sides printed with a black identification
code.
Complies.

Thickness: Mean of 10: 1.50-1.65 mm.
Complies.

Performance:

Test strains:

Candida albicans ATCC 10231
Candida pseudotropicalis MIHE 2337

Product complies with specifications.

Uniformity: Inhibition zones of individual tablets
vary less than ± 1 mm from mean.
Complies.

Microbiological quality: Complies.

Rosco Diagnostica A/S



Peter Samuelsen
Manager Quality Assurance