

## QUALITY CONTROL CERTIFICATE

### ECONAZOLE 10 µg Neo-Sensitabs

REF: 81312

LOT: 1902-1

Date of analysis: 2019-12

Date of expiry: 2024-12

Appearance: Light yellow, plane tablet, 9 mm in diameter,  
on both sides printed with a black identification  
code.  
Complies.

Thickness: Mean of 10: 1.50-1.65 mm.  
Complies.

Performance:

Test strains:

*Candida albicans* ATCC 90028

*Candida krusei* ATCC 6258

Product complies with specifications.

Uniformity: Inhibition zones of individual tablets  
vary less than  $\pm 1$  mm from mean.  
Complies.

Microbiological quality: Complies.

ROSCO DIAGNOSTICA A/S



Peter Samuelson  
Quality Assurance Manager