

QUALITY CONTROL CERTIFICATE

MICONAZOLE 10 µg Neo-Sensitabs

REF: 82012

LOT: 1601-2

Date of analysis: 2016-03

Date of expiry: 2021-03

Appearance: Light yellow, plane tablet, 9 mm in diameter,
on both sides printed with a black identification
code.
Complies.

Thickness: Mean of 10: 1.50-1.65 mm.
Complies.

Performance:

Test strains:


Candida albicans ATCC 90028
Candida parapsilosis ATCC 22019
Candida krusei ATCC 6258

Product complies with specifications.

Uniformity: Inhibition zones of individual tablets
vary less than ± 1 mm from mean.
Complies.

Microbiological quality: Complies.

Rosco Diagnostica A/S



Michela Venturi
Manager Quality Assurance