

QUALITY CONTROL CERTIFICATE

VORICONAZOLE 1 µg Neo-Sensitabs

REF: 82312N

LOT: 1501-1

Date of analysis: 2015-03

Date of expiry: 2020-03

Appearance: Light yellow, plane tablet, 9 mm in diameter,
on both sides printed with a black identification
code.
Complies.

Thickness: Mean of 10: 1.50-1.65 mm.
Complies.

Performance:

Test strains:


Candida albicans ATCC 90028
Candida parapsilesis ATCC 22019
Candida krusei ATCC 6258

Product complies with specifications.

Uniformity: Inhibition zones of individual tablets
vary less than ± 1 mm from mean.
Complies.

Microbiological quality: Complies.

Rosco Diagnostica a/s



Dorrit Reumert

Quality Assurance Manager