

## QUALITY CONTROL CERTIFICATE

### **VORICONAZOLE 1 µg Neo-Sensitabs**

REF: 82312

LOT: 1701-4

Date of analysis: 2017-04

Date of expiry: 2022-03

Appearance: Light yellow, plane tablet, 9 mm in diameter,  
on both sides printed with a black identification  
code.  
Complies.

Thickness: Mean of 10: 1.50-1.65 mm.  
Complies.

Performance:

Test strains:


*Candida albicans* ATCC 90028  
*Candida parapsilesis* ATCC 22019  
*Candida krusei* ATCC 6258

Product complies with specifications.

Uniformity: Inhibition zones of individual tablets  
vary less than  $\pm 1$  mm from mean.  
Complies.

Microbiological quality: Complies.

Rosco Diagnostica A/S



Michela Venturi  
Manager Quality Assurance