

QUALITY CONTROL CERTIFICATE

FLUCONAZOLE 25 µg Neo-Sensitabs

REF: 82512

LOT: 210039

Date of analysis: 2021-10

Date of expiry: 2026-09

Appearance: Light yellow, plane tablet, 9 mm in diameter,
on both sides printed with a black identification
code.
Complies.

Thickness: Mean of 10: 1.50-1.65 mm.
Complies.

Performance:

Test strains:

Candida Paraposilosis ATCC 22019
Candida albicans ATCC 90028

Product complies with specifications.

Uniformity: Inhibition zones of individual tablets
vary less than ± 1 mm from mean.
Complies.

Microbiological quality: Complies.

Rosco Diagnostica a/s



Peter Samuelson
Quality Assurance Manager