

# • Certificate of Analysis

Product Name : **Bruker Bacterial Test Standard**

Part No. **8255343**  
 LOT No.: **0000268920**  
 Use by<sup>1)</sup>: **2018-02**  
 Date of Production: **2017-03-16**


Storage Conditions : **Store below -18 °C**  
**Stability of dissolved BTS: 5 months below -18°C**

<u>Test</u>	<u>Specification</u>	<u>LOT Results</u>
Mass accuracy of reference proteins by MALDI-TOF MS	Within a mass range of $\pm 600$ ppm	Passed
Relative intensity of protein No. 3 and No. 4 by MALDI-TOF MS	$0.25 \pm 0.06$	Passed
Side chain reaction	Less than 20% according to MALDI-TOF analysis	Passed

**Remarks:**

The BTS is shipped at ambient temperatures. We recommend storing the standard below -18°C on arrival. Dissolved BTS standard should be aliquoted and frozen. For usage, thaw the aliquots at room temperature and mix the content of the micro tube by snipping with a finger to the tube. The remaining dissolved BTS should be stored as soon as possible below -18°C.

<sup>1)</sup> if handled as advised in the instruction for use.



**i.V. Jens Löffelholz**  
 Quality Assurance Manager

Bruker Daltonik GmbH, Germany  
 Fahrenheitstr. 4 – 28359 Bremen  
 phone +49-421-2205-0

QC release date: 17/03/2017

Not for use in diagnostic procedures. For research use only.