

CERTIFICATE OF ANALYSIS

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and testing specifications required by Cepheid's Quality System, in compliance with the US Food and Drug Administration's Quality System Requirements, ISO 13485, European IVD Directive and the Canadian Medical Device Regulations (CMDR).

Product Name: Xper	rt® CT/NG			
Cepheid Catalogue I	Part No.: GXCT/NG-CE-10			
Kit Lot No.: 1000	0144573			
Cartridge Lot No.:31	1106			
Kit Expiration Date:	2021-01-17			
Legal Manufacturer		Manufacturing Facility	Solna (Sunnyvale
Cepheid AB		Cepheid AB		dente de la constitución de la c
Röntgenvägen 5 SE-17154 Solna		Röntgenvägen 5 SE-171 54 Solna		
Sweden Sweden		Sweden Sweden		
			1	
Functional Testing				
Test Description	Acceptance Criteria			Test Result

Test Description	Acceptance Criteria	Test Result
Positive	CT detected; NG detected	Passed
Negative	CT not detected; NG not detected	Passed

Alcold 2	10190211
Signature of Quality Assurance	Date
Name: Samaneh Vahid	
Title: OA Analyst	